

**CarnoCheck®**

## Instructions For Use

Parallel qualitative detection of 8 animal species constituents in meat food products

REF 462 030

For professional users only

Revision: BQ-020-02  
May 2015



**Greiner Bio-One GmbH**  
Maybachstr. 2 • 72636 Frickenhausen • Germany  
Phone: +49 (0) 7022 948-0 • Fax +49 (0) 7022 948-514  
info@de.gbo.com • www.gbo.com/bioscience

  
**greiner bio-one**











DANGER



30



## GLOSSARY OF SYMBOLS

											
en	Store in the dark	Use by	Consult Instructions For Use	Catalog Number	Manufacturer	Temperature limitation	Contents sufficient for <n> tests	Warning	Danger	Batch code	Important Note
de	Im Dunkeln lagern	Mindestens haltbar bis	Vor Gebrauch Anweisung lesen	Katalognummer	Hersteller	Temperaturbegrenzung	Inhalt ausreichend für <n> Tests	Achtung	Gefahr	Chargenbezeichnung	Wichtiger Hinweis
fr	À stocker à l'abri de la lumière	Date limite de conservation jusqu'au	Lire les instructions avant utilisation	Numéro de référence	Fabricant	Limite de température	Contenu suffisant pour <n> tests	Attention	Danger	Code du lot	Note importante
es	Conservar en un lugar oscuro	A utilizar preferiblemente antes de	Antes de usar, lea las instrucciones	Número de catálogo	Fabricante	Limitación de temperatura	Contenido suficiente para <n> ensayos	Atención	Peligro	Código de lote	Nota importante
it	Conservare al buio	Da utilizzare entro e non oltre	Leggere le istruzioni prima dell'uso	Numero catalogo	Produttore	Limitazione temperatura	Contenuto sufficiente per test <n>	Attenzione	Pericolo	Codice del lotto	Nota importante
pt	Conservar num local escuro	A utilizar preferivelmente antes de	Antes de usar, leia as instruções	Número de catálogo	Fabricante	Limitação de temperatura	Conteúdo suficiente para <n> ensaios	Atenção	Perigo	Código do lote	Aviso importante
nl	Donker bewaren	Tenminste houdbaar tot	Gebruiksaanwijzing lezen	Catalogusnummer	Fabrikant	Temperatuurbepending	Voldoende inhoud voor <n> tests	Waarschuwing	Gevaar	Lot nummer	Belangrijke opmerking
da	Opbevares mørkt	Anvendes senest	Læs brugsanvisningen	Katalognummer	Producent	Temperaturbegrænsning	Indeholder nok til <n> test	Advarsel	Fare	Lotnummer	Vigtig henvisning
sv	Förvaras mörkt.	Sista förbrukningsdag	Läs bruksanvisningen före användning	Katalognummer	Tillverkare	Temperaturbegränsning	Innehållet räcker till <n> tester	Varning	Fara	Lot nummer	Viktigt meddelande
pl	Przechowywać w ciemności	Termin zdatności	Przed użyciem przeczytać instrukcję	Numer katalogowy	Producent	Ograniczenie temperatury	Zawartość wystarcza na <n> testów	Uwaga	NIEBEZPIECZESTWO	Kod partii	Ważne
no	Oppbevares mørkt	holdbar til	Les bruksanvisning før bruk	katalognummer	produsent	temperaturbegrensning	Innhold tilstrekkelig for <n> tester	Obs	Fare	batch nr.	Viktig merknad
el	Αποθηκεύεται στα σκοτεινά	το λιγότερο διατηρείται	πριν την χρήση διαβάστε τις οδηγίες	Αριθμός Καταλόγου	Παραγωγός	περιορισμός θερμοκρασίας	Περιεχόμενο αρκετό για <n> τεστ	ΠΡΟΣΟΧΗ	ΚΙΝΔΥΝΟΣ	κωδικός παρτίδας	Σημαντική υπόδειξη
tr	Karanlık yerde saklayınız	Son kullanma tarihi:	Kullanmadan önce talimatı okuyun	Katalog numarası	Üretici firma	Sıcaklık sınırlaması	İçeriği <n> test için yeterlidir	Dikkat	Tehlikeli	Partikodu	Önemli Not

# TABLE OF CONTENTS

<b>1. KIT CONTENTS .....</b>	<b>5</b>
<b>2. CONSUMABLES, EQUIPMENT AND HARDWARE REQUIRED .....</b>	<b>6</b>
<b>3. SHIPMENT AND STORAGE.....</b>	<b>8</b>
<b>4. SAFETY INSTRUCTIONS .....</b>	<b>8</b>
<b>5. WASTE DISPOSAL .....</b>	<b>10</b>
<b>6. INTRODUCTION .....</b>	<b>11</b>
6.1 Intended use .....	12
6.2 Assay principle.....	13
6.3 Design of the CarnoCheck® DNA chip .....	15
6.3.1 CarnoCheck® chip layout .....	15
6.3.2 On-chip controls.....	16
<b>7. INSTRUCTIONS FOR THE CARNOCHECK® WORKFLOW .....</b>	<b>17</b>
7.1 General instructions .....	17
7.2 Room separation.....	17
7.3 Warnings and precautions .....	18
7.3.1 Contamination prevention .....	18
7.3.2 Instruction for handling DNA chips.....	18
7.3.3 General precautions.....	18
7.3.4 Working safely .....	19
<b>8. CARNOCHECK® PROCEDURE .....</b>	<b>20</b>
8.1 Sample collection, homogenisation and DNA extraction.....	20
8.1.1 Sample collection.....	20
8.1.2 Homogenisation .....	20
8.1.3 DNA extraction .....	22
8.2 Polymerase chain reaction (PCR).....	23
8.2.1 Thermal cycler set-up .....	23
8.2.2 PCR reaction set-up.....	24
8.3 Hybridisation and washing .....	26
8.3.1 Preparation and set-up .....	26
8.3.2 Hybridisation .....	28
8.3.3 Washing and drying .....	30
8.4 Scanning and evaluation of the CarnoCheck® chip.....	32
<b>9. TROUBLESHOOTING .....</b>	<b>33</b>

<b>10. TECHNICAL ASSISTANCE .....</b>	<b>34</b>
<b>11. PERFORMANCE CHARACTERISTICS OF CARNOCHECK® .....</b>	<b>35</b>
11.4.1 Procedure .....	36
11.4.2 Results .....	36
11.4.3 Successful participation at FAPAS® Proficiency Test .....	38
<b>12. CARNOCHECK® SHORT PROTOCOL .....</b>	<b>39</b>
12.1 Room 2: PCR - Set-up of reaction mix .....	39
12.2 Room 2: PCR - DNA template addition / PCR reaction .....	40
12.3 Room 3: Hybridisation - Preparation / Hybridisation reaction.....	41
12.4 Room 3: Hybridisation - Preparation / Hybridisation reaction.....	42

## 1. KIT CONTENTS

Components	Description	Quantity
CarnoCheck® 5x6 Arrays	Slidebox with 5 CarnoCheck® chips imprinted with 6 CarnoCheck® arrays in wells A1-A6	5 x 6 arrays
CarnoCheck® PCR MasterMix	PCR MasterMix contains all components required for PCR except Taq Polymerase	2 x 450 µl
CarnoCheck® Hyb.-Buffer	Hybridization Buffer	1000 µl
CarnoCheck® Buffer A, conc.	concentrate of Buffer A	40 ml
CarnoCheck® Buffer B, conc.	concentrate of Buffer B	15 ml
Flyer	Download instructions for Instructions For Use	1

## 2. CONSUMABLES, EQUIPMENT AND HARDWARE REQUIRED

Consumables	Greiner Bio-One Cat. No.	Quantity (pieces)
CarnoCheck® test kit	462 030	30 reactions
<b>Sterile, DNase-free micropipette filter tips</b> <sup>1</sup>		
0.5-10 µl filter tips	765 288	960
0.5-20 µl filter tips	774 288	960
10-100 µl filter tips	772 288	960
10-200 µl filter tips	739 288	960
100-1000 µl filter tips	750 288	600
<b>DNase-free reaction tubes</b>		
Reaction tube 1.5 ml	616 201	4000
Reaction tube 0.2 ml <sup>2</sup>	683 201	1000
8 x 0.2 ml PCR strips	673 210	1250
Cap strips for 8 x 0.2 ml PCR strips	373 270	1250
<b>Plastic pipettes for pipettor</b>		
Pipette 10 ml	607 180 or 607 160	200
Pipette 50 ml	768 180 or 768 160	100
50 ml polypropylene tubes (optional) <sup>3</sup>	210 261	450

<sup>1</sup> Some of the mentioned tip sizes are optional depending on the micropipettes available.

<sup>2</sup> In principle, it is recommended to use 8-tube PCR strips. Single reaction tubes (0.2 ml) are optional if strips are not available.

<sup>3</sup> Only required if a slide centrifuge is not available.

Equipment	Greiner Bio-One Cat. No.	Quantity
CheckScanner™	862 070	1
CheckReport™Software Basic	862 080	1
CheckReport™Software CarnoCheck® plugin	862 084	1
oCheck® Hybridisation Chamber with slideholder	447 070	1
Handle for slideholder	447 001	1
oCheck® Washbox	447 020	2
<b>Enzymes required</b>		
• AmpliTaq® Gold DNA Polymerase (Applied Biosystems, Cat No. N808-0240, 5 U/µl, 250 U)		
<b>DNA Extraction Kits</b>		
• NucleoSpin® Food (Macherey-Nagel, Cat. No. 740945.50, for 50 purifications)		
• Gen Elute™ (Sigma-Aldrich, Cat. No. G1N-50, for 50 purifications)		
<b>Additional consumables required</b>		
<ul style="list-style-type: none"> <li>• PCR-grade water</li> <li>• ethanol puriss. p.a., ≥ 99,8 %</li> <li>• distilled or deionised water</li> <li>• single-use gloves</li> <li>• cleaning solution for DNA decontamination</li> <li>• scalpel blades</li> <li>• disposables for homogenization (depending on the homogenization procedure)</li> </ul>		

#### Additional equipment required

- Precision scales
- Scalpel with disposable blades
- Equipment for homogenisation
- Clean bench for PCR set up
- Cooling block for Taq
- Microcentrifuge for 1.5 and 2 ml reaction tubes
- Centrifuge for 50 ml polypropylene tubes (e.g. BeckmanCoulter; Allegra X-22 Centrifuge; C0650 Fixed-Angle Rotor) or slide centrifuge (e.g. Labnet: Slide Spinner; VWR International: Galaxy MiniArray Centrifuge)
- Microcentrifuge for single 0.2 ml reaction tubes or 8-tube PCR strips (e.g. Labnet: Spectrafuge Mini Centrifuge)
- PCR thermal cycler:
  - GeneAmp® PCR system 9700 (Applied Biosystems) or Veriti™ 96-Well Thermal Cycler (Applied Biosystems)
- Water bath or heating block (65 °C and 70 °C)
- Micropipettes (different ranges from 1 - 1000 µl)
- 8-Channel multipipette (range: 5 - 50 µl), e.g. Brand Transferpette®-8 (Brand)
- Pipettor for glass and plastic pipettes
- Vortex shaker
- Racks for different reaction tubes
- Waste container
- Timer
- Photometer for determination of DNA concentration

#### Additional hardware required

- Computer (for system requirements see Instructions For Use of the CheckSanner™ and the CheckReport™Software)



### 3. SHIPMENT AND STORAGE

The shipment of the CarnoCheck® test kit takes place at room temperature. Nevertheless, the kit has to be stored immediately upon receipt **at 4-8 °C** and should be protected from light. Stored correctly, the CarnoCheck® test kit and its components can be used until the indicated expiration date. Furthermore, under these conditions the shelf life does not deviate from the expiration date after the first opening of the kit and its components.

### 4. SAFETY INSTRUCTIONS

The CarnoCheck® test kit is for laboratory use only, not for drug, household, or other purposes. Always wear a suitable lab coat, disposable gloves, and protective goggles and follow the safety instructions given in this section.

The following components of the CarnoCheck® test kit contain harmful or hazardous contents.

Kit Component	Hazardous content (Quantity)	Classification according to regulation (EC) No 1272/2008	GHS pictogram and signal word	Hazard and precautionary statements	
CarnoCheck® Hybridisation Buffer	Guanidine thiocyanate (20-60%)  CAS No. 593-84-0	acute toxicity, oral, (category 4),	 <b>DANGER</b>	H302	Harmful if swallowed
				H332	Harmful if inhaled
		acute toxicity, inhalation, (category 4),		H314	Causes severe skin burns and eye damage
		chronic aquatic toxicity, (category 3)		H412	Harmful to aquatic life with long lasting effects
		skin corrosive (category 1c)		P273	Avoid release to the environment
				P280	Wear protective gloves/protective clothing/eye protection/face protection
				P305+P351+P338	IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. Immediately call a POISON CENTER or doctor/physician
CarnoCheck® Hybridisation Buffer	Sodium dodecyl sulfate (< 20%)  CAS No. 151-21-3	skin irritation (category 2),	 <b>DANGER</b>		Supplemental hazard information (EU): EUH032
		serious eye damage, (category 1)			
				H315	Causes skin irritations
				H318	Causes serious eye damage
				P280	Wear protective gloves/eye protection/face protection
				P305+P351+P338	IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing



The current version of the Safety Data Sheet for this product can be downloaded from the Greiner Bio-One website: [www.gbo.com/bioscience/biochips\\_download](http://www.gbo.com/bioscience/biochips_download)

## 5. WASTE DISPOSAL

After washing and drying of the CarnoCheck® chip, the washing solutions I and II can be discarded without any special precautions. Dispose the used CarnoCheck® chip, unused kit components as well as unused hybridisation mix with the laboratory chemical waste. Observe all national, state, and local regulations regarding disposal.

## 6. INTRODUCTION

**CarnoCheck®** is a DNA-Chip for the parallel qualitative detection of eight different animal species constituents in meat containing food products. With **CarnoCheck®** the animal species can be rapidly and simultaneously identified in raw and preserved food matrices derived from meat.

Preserved foods differ in the extent to which they have been heated (and thereby in their shelf life). The extent of preservation is defined by the sterilization value F. The F value for a sterilisation process is the number of minutes required to kill a known population of microorganisms in a given food under specified conditions. When F is used without a subscript indicating temperature, 121,1 °C is assumed, so an F-value of 1 corresponds to 1 minute at 121 °C.<sup>1</sup>

- Cooker preserves / boiled food: Heating is at temperatures up to a maximum of 100 °C. The shelf life when kept cooled at temperatures lower than 10 °C is limited to 1 year.  $F \geq 0,4$ .
- Fully preserved food: Heating is at temperatures over 100 °C. The shelf life is limited to 4 years at 25 °C without cooling.  $F \geq 3$ .
- Tropical preserves: Heating is at temperatures over 100 °C. The shelf life is limited to 1 year at 40 °C without cooling.  $F \geq 12$ .

Cooked and full preserves comprise approximately 90% of all preserved foods.

Raw, cooked and full preserves can be analysed with **CarnoCheck®**, but tropical preserves cannot. The extreme heating in the making of tropical preserves breaks the DNA into very small fragments. As a result, amplification by PCR for **CarnoCheck®** analysis is not possible.

**CarnoCheck®** is validated for meat, mixtures of meat and sausages with a sterilisation value of  $F \leq 3,4$ . The low limit of detection (0.05- 0,35% depending on species, processing of the food product and purification) permits the detection of even small traces of animal constituents.

<sup>1</sup> For complete definition of sterilisation symbols (D, Z, F) refer to <http://www.fda.gov/ICECI/Inspections/InspectionGuides/ucm070835.htm>

## 6.1 Intended use

CarnoCheck® is an analysis kit for the detection of pig, cattle, sheep, turkey, chicken/(duck)<sup>2</sup>, horse, donkey, goat in raw, cooker preserved, boiled or fully preserved meat, mixtures of meat and sausages with a sterilization value of  $F \leq 3,4$ .

**Table 1: Matrices, preservation grades and animal species analysable with CarnoCheck®**

matrices	preservation grade	species
meat mixtures of meat sausages <sup>3</sup>	raw cooker preserved boiled fully preserved $F \leq 3,4$	pig cattle sheep turkey chicken <sup>2</sup> horse donkey goat

<sup>2</sup> In case of detection of a chicken species signal, the presence of duck species cannot be excluded. Not all duck species give a positive signal on the chicken probe, detection of duck species is not the intended use of CarnoCheck®.

<sup>3</sup> Composition of the tested sausages: 50% meat, 25% plant oil, 23% ice, 1,5% , nitrit pickling salt, 0,25 % spices (pepper, muscat, cardamon, cinnamon).

Since the limit of detection of CarnoCheck® depends on the species, composition, processing and preservation grade, the testing of complex food products should be validated first by verification of the analysis procedure by reference samples with known species content.

## 6.2 Assay principle

CarnoCheck® is a microarray-based test kit for the detection of species specific differences in the sequence of the cytochrome b (cyt b) gene (see **Table 2**). The assay procedure is summarised in Figure 1. Prior to the CarnoCheck® analysis, DNA must be extracted from a meat sample.

Sample collection, homogenisation and DNA extraction are not part of the CarnoCheck® test kit. After the extraction of genomic DNA from a meat sample, a 389 bp fragment of cyt b gene of all the animal species present in the sample is amplified and fluorescently labeled with the aid of the polymerase chain reaction. This is achieved using a primer pair matching two highly conserved sequences, such that only a single reaction is necessary.


In the same reaction, an internal control-template present in the CarnoCheck® PCR MasterMix is amplified to monitor the performance of the PCR (PCR control). The PCR products are then hybridised to specific DNA probes and on-chip controls attached to the CarnoCheck® chip surface. Every chip contains 6 DNA-microarrays (in wells A1-A6), allowing the simultaneous analysis of 6 meat samples. Unbound DNA is removed in the subsequent washing steps. The hybridisation efficiency is monitored (hybridisation control). Finally, the CarnoCheck® chip is automatically scanned, analysed and evaluated using the CheckScanner™ and CheckReport™ Software, respectively (see ordering information in Chapter 2). The CheckScanner™ is a two colour laser scanner (excitation wavelengths of 532 nm and 635 nm), which enables the detection of the fluorescence signal generated by the presence of cyt b-specific amplification products as well as the controls (see chapter 6.3.2). The CheckReport™ Software allows the visualisation, analysis and evaluation of the results and automatically shows the corresponding values of both the detected animal species and the controls in a detailed report. The report clearly indicates the presence or absence of one or more of the 8 animal species detectable and the comprehensive on-chip controls render the analysis highly reliable.

**Table 2: The test is based on species specific differences of cyt b genes from Locus/Accession Numbers:**


LOCUS/ ACCESSION*	ORGANISM	SPECIES
DQ186273	Bos taurus	cattle
DQ089479	Capra hircus	goat
X97337	Equus asinus	donkey
DQ223538	Equus caballus	horse
X52392	Gallus gallus	chicken
L08381	Meleagris gallopavo	turkey
DQ097414	Ovis aries	sheep
AY534296	Sus scrofa	pig

\*Source: <http://www.ncbi.nlm.nih.gov/nuccore>


### 1. PCR reaction




### 2. Hybridisation



### 3. Washing and drying



### 4. Scanning and analysis




**Figure 1: CarnoCheck® assay procedure**

- 1. PCR reaction:** After DNA extraction, a 389 bp fragment of the cyt B gene and a control target are amplified by PCR and fluorescently labeled. The amplification products are then hybridised to complementary DNA probes on the chip.
- 2. Hybridisation:** Each species is detected by a specific DNA probe present in five replicates.
- 3. Washing & drying:** Unbound DNA is removed in the subsequent washing steps.
- 4. Scanning & analysis:** The CarnoCheck® chip is scanned, analysed and evaluated using the CheckScanner™ and CheckReport™ Software. A report is created that clearly indicates the presence or absence of one or more of the animal species detectable.

## 6.3 Design of the CarnoCheck® DNA chip

### 6.3.1 CarnoCheck® chip layout

Each CarnoCheck® chip contains 6 microarrays designated as well A1 - A6. Each CarnoCheck® microarray comprises 12 different probes and is bordered by an elevated rim. Each probe is spotted in five replicates. The CarnoCheck® microarray layout is illustrated in Figure 2 and the on-chip controls are further explained in Chapter 6.3.2.



**Figure 2: Design of the CarnoCheck® chip**

**a)** Schematic drawing of the CarnoCheck® chip. Each CarnoCheck® chip contains 6 microarrays designated as well A1 - A6. **b)** and **c)** Images displayed by the CheckReport™ Software for the two different excitation wavelengths used for scanning (**b)** red channel: 635 nm; **c)** green channel: 532 nm) and schematic drawings of the CarnoCheck® microarray layout. Species-specific probes and on-chip controls are indicated.

### 6.3.2 On-chip controls

The design of the **CarnoCheck®** DNA-chip incorporates comprehensive on-chip controls. Several control systems monitor critical steps of both the assay and chip processing, including quality of the PCR reaction (PCR control), the efficiency of the hybridisation (hybridisation control), as well as spot homogeneity and printing quality (orientation control and printing control). The CheckReport™ Software automatically shows both the corresponding values of the controls and the detected animal species in a detailed report. For read-out of the different controls, both excitation wavelengths of the CheckScanner™ are used. For the control of PCR performance (PCR control), the red channel is used (excitation wavelength of 635 nm), while the quality of the hybridisation and the chip (hybridisation, orientation and printing control) is assessed in the green channel (excitation wavelength of 532 nm).

#### PCR control

**CarnoCheck®** monitors the quality of the PCR reaction. Amplification of an internal control template present in the **CarnoCheck®** PCR MasterMix generates a signal on the PCR control spots on the **CarnoCheck®** chip. The quality of the amplification reaction is also automatically assessed by the CheckReport™ Software. If PCR performance is below a predefined threshold, the CheckReport™ Software will indicate the PCR control as “failed” and the analysis must be repeated (see Chapter 9).

If the amount of animal specific DNA in the sample is very high, the fluorescence signal of the PCR control spots may be low or even absent due to competition during the PCR reaction. In this case, the fluorescence signal for at least one animal-specific probe must exceed a predefined threshold in order for the test to be considered valid.

#### Hybridisation control

**CarnoCheck®** monitors the efficiency of the hybridisation through use of a fluorescence labeled probe within the **CarnoCheck®** Hybridisation Buffer, which hybridises to specific DNA sequences on the **CarnoCheck®** chip. An adequate hybridisation efficiency results in fluorescence signals on each array spot. The results of five hybridisation control spots on the **CarnoCheck®** chip are also assessed by the CheckReport™ Software.

#### Orientation and printing control

The orientation control spots of the **CarnoCheck®** chip generate fluorescence signals irrespective of the efficiency of the hybridisation process. These spots are used by the CheckReport™ Software as guidance points for a correct spot finding, which is a prerequisite for the correct analysis of the signals. In addition, the quality of the printing process is monitored by the presence of a green fluorescence signal at each chip spot (printing control).



## 7. INSTRUCTIONS FOR THE CARNOCHECK® WORKFLOW


### 7.1 General instructions

When implementing currently used state-of-the-art techniques in molecular biology into a laboratory, the following instructions must be considered to ensure both maximum safety for laboratory staff and high quality results.

Execution of molecular biology techniques such as DNA extraction, amplification and detection of the amplification products require appropriately qualified personnel. In addition, a clean and well structured workflow is required to prevent erroneous results, such as those occurring due to DNA degradation or contamination by amplification products. To ensure this, it is necessary to separate the areas of extraction, amplification and detection as described in Chapter 7.2. Each area should be equipped with separate equipment, consumables, lab coats and gloves. Never transfer lab coats, gloves or equipment from one distinct area to another.

### 7.2 Room separation

**Figure 3** shows an example of how a laboratory may be separated into three distinct sections. One is used only for DNA extraction, another is for the set-up and running of PCR reactions and the last is for hybridisation and analysis. Each room is used exclusively for the application or technique indicated to prevent sample contamination. The use of colour coding could be advantageous to avoid the accidental exchange of equipment and consumables between areas.



**Figure 3: Room separation**

**Room 1:** The entire DNA extraction procedure must be performed in this room.

**Room 2:** Within this room, the reaction mix for the PCR is set up and aliquoted (optimally under a PCR hood). The addition of the DNA samples extracted in room 1 must be carried out in a separate space within room 2.

**Room3:** Within the third laboratory room the hybridisation reaction, washing steps and chip drying take place. Additionally, the CheckScanner™ in conjunction with the CheckReport™Software is used for the final analysis of the CarnoCheck® assay.



Neither equipment nor consumables should be interchanged between the different laboratory rooms and spaces. Hence, duplications in equipment and consumables are a necessity and should be taken into account when equipping the laboratory.

## 7.3 Warnings and precautions

### 7.3.1 Contamination prevention

- Lab coats must be worn throughout the procedure and different sets of lab coats are required for each laboratory room.
- Gloves must be worn during each step of the analysis and must be frequently changed, especially during DNA extraction.
- The working place must be decontaminated with an appropriate cleaning solution.
- Never touch the inside of a reaction tube cap. To avoid cross-contamination, open only one tube at a time.
- Appropriate micropipette filter tips with aerosol barrier must be used (free of DNase, RNase and human DNA). Pipette tips should always be changed between liquid transfers.

### 7.3.2 Instruction for handling DNA chips

- DNA chips should be used in a dust-free environment. The deposition of dust and other particles on the chip surface must be prevented.
- Do not touch the hybridisation zone on the chip surface.
- Only the labelled side of the chip is intended for hybridisation.
- Do not use any marker pens for the identification of DNA chips, as they lead to unspecific fluorescence on the chip.
- DNA arrays are for single use only. Hybridised chips cannot be reused.
- Store unused chips in the original box inside the delivered zipper bag containing the desiccant.

### 7.3.3 General precautions

- This kit is exclusively for the qualitative detection of animal constituents in meat products and should only be used by personnel trained in molecular biology laboratory practice.
- Upon arrival, check the kit components for damage. If one of the components is damaged (e.g. buffer bottles), contact your local Greiner Bio-One distributor. Do not use damaged kit components, as their use may lead to poor kit performance.
- Do not use the **CarnoCheck®** test kit after the expiry date.
- Do not use expired reagents.
- Do not mix reagents from different batches.
- Use only reagents/equipment provided with the kit and those recommended by the manufacturer.
- Regular calibration/maintenance should be performed for micropipettes, waterbath and heating block.
- Pipetting of small amounts of liquid in the microliter range is a challenge. Therefore take care to pipette as accurately as possible.
- To avoid microbial contamination of the reagents, take care when removing aliquots from reagent tubes.
- All centrifugation steps should be carried out at room temperature (18-25 °C).

#### 7.3.4 Working safely

- Never pipette solutions by mouth.
- Do not eat, drink, smoke or apply cosmetic products in the work areas.
- Avoid direct contact with the biological samples as well as splashing or spraying of the samples.
- Always wear a lab coat, gloves and goggles while working with human samples.
- Wash hands carefully after handling of samples and reagents.

## 8. CARNOCHECK® PROCEDURE

The following chapter describes the different working steps which finally lead to the production of a detailed report indicating the presence or absence of one or more of the 8 animal species detectable in the analysed meat sample. **Figure 4** shows an overview of the necessary different working steps. It also indicates the corresponding subchapter describing the specific assay step. The working steps must be performed in the order outlined in this chapter. Each specific hands-on step is indicated by a blue arrow ➡.



Sample collection, homogenisation, DNA extraction and analysis with the CheckReport™ Software are not part of the CarnoCheck® test kit. Therefore, the description of these working steps is abbreviated within this chapter. For more detailed information, please refer to the corresponding Instructions For Use, e.g. from the DNA Extraction Kits and the CheckReport™ Software.

### 8.1 Sample collection, homogenisation and DNA extraction


#### 8.1.1 Sample collection

For sample collection knives or scalpels with sterile disposable blades are recommended. In order to avoid cross contamination of samples use for each sample a new disposable blade.


#### 8.1.2 Homogenisation

If a homogenisation step is necessary, depends on the homogeneity of the starting material. For inhomogeneous samples representative random samples should be taken at several sites and combined in a single homogenate. In a homogenous sample (e.g. raw meat) the excision of one representative aliquot of the sample and direct lysis of this aliquot without homogenisation is sufficient. Commercial homogenisers, for example laboratory blenders such as the stomacher® Circulator or bead mills, can be used. Avoid cross contamination of samples by using disposable homogenisation equipment.

### Chapter 8.1.2 DNA extraction



### Chapter 8.2 PCR



### Chapter 8.3.2 Hybridisation



### Chapter 8.3.3 Washing & drying



### Chapter 8.4 Scanning & evaluation




Figure 4: Overview of the different CarnoCheck® working steps

### 8.1.3 DNA extraction

Within the oCheck® workflow, DNA extraction is the process most sensitive to contamination and so strict adherence to the contamination prevention guidelines outlined in chapter 7 is essential.

CarnoCheck® has been validated using DNA prepared with the following DNA extraction kit and protocol:

- NucleoSpin® Food (Macherey-Nagel, Düren, Germany) User Manual - Protocol for genomic DNA purification from food

Please follow the protocol provided by the DNA extraction kit's manufacturer. Start the DNA extraction from 100 mg of a homogenous sample. Depending on the individual sample, complete lysis of the sample can take 1-3 hours or overnight. If the sample is not completely lysed after a long incubation time, take off the supernatant for DNA extraction and discard the remaining unlysed sample. After DNA extraction we recommend a determination of the concentration of the DNA. The concentration of DNA-template for the PCR should be between 5 - 15 ng/µl. The detection limit of all species is below 0.35% in a mixed sample. However, the detection limit of CarnoCheck® – depends on the extent to which the sample has been processed and the purity of the DNA, i.e. the presence of possible PCR inhibitors.



It is recommended to include a negative extraction control for every time a DNA extraction is performed. As negative extraction control sample, the DNA elution buffer of the appropriate DNA extraction kit or PCR-grade water may be used. If a homogenisation of the tested samples is performed, perform also the homogenisation procedure with this negative extraction control sample in order to monitor carry-over contamination in the homogenisation and DNA extraction procedure.

## 8.2 Polymerase chain reaction (PCR)

PCR is a very sensitive method which can detect extremely small amounts of DNA. Special precautions must be observed in order to avoid reaction contamination (see Chapter 7). AmpliTaq Gold® DNA Polymerase (Applied Biosystems, 5U/μl) is required but not provided with the CarnoCheck® test kit and must be purchased separately (see Chapter 2).



The CarnoCheck® test kit has been validated using AmpliTaq Gold® DNA Polymerase from Applied Biosystems (see ordering information in Chapter 2). It is mandatory to use this enzyme in order to achieve the established performance.

### 8.2.1 Thermal cyclers set-up

The CarnoCheck® test kit has been validated with the following thermal cyclers:

- GeneAmp® PCR system 9700 (Applied Biosystems)
- Veriti™ 96-Well Thermal Cycler (Applied Biosystems).



It is absolutely necessary to use one of the thermal cyclers mentioned above in order to achieve the established performance.

The thermal cycler program of the CarnoCheck® PCR is summarised in **Table 2**.

**Table 2: Thermal cycler program of the CarnoCheck® PCR**

Time	Temp. °C	No. of cycles
10 min	94 °C	1
30 s	94 °C	40
30 s	53 °C	
30 s	72 °C	
5 min	72 °C	1
Hold	10 °C	

In addition, the following run parameters must be set for each thermal cycler. For a description on how to set these parameters see the Instructions For Use of the respective thermal cycler.

#### **GeneAmp® PCR system 9700 (Applied Biosystems)**

Set the reaction volume to 25 μl, the ramp speed to “9600” and use lid temperature of 103 °C.

#### **Veriti™ 96-Well Thermal Cycler (Applied Biosystems)**

Use the Convert Method tool of the Veriti™ 96-Well Thermal Cycler to enter the CarnoCheck® PCR program and choose the “9600 Emulation Mode”. Set the reaction volume to 25 μl and the temperature of the lid to 103 °C.

## 8.2.2 PCR reaction set-up

With the exception of the AmpliTaq Gold® DNA Polymerase the CarnoCheck® PCR MasterMix already contains all components necessary for performing the PCR reaction (PCR buffer, MgCl<sub>2</sub>, dNTPs, primers, PCR control template).

The set-up of the reaction mix is optimally performed in a protected surrounding, e.g. a PCR hood, to avoid reaction contamination.

- ➡ Prepare the reaction mix (consisting of CarnoCheck® PCR MasterMix, AmpliTaq Gold® DNA Polymerase for the required quantity of PCR reactions as outlined in **Table 3**.

**Table 3: Set-up of the CarnoCheck® PCR reaction**

	1 reaction	7 reactions (1 chip)
CarnoCheck® PCR MasterMix	23.8 µl	166.6 µl
AmpliTaq Gold® DNA Polymerase (5 U /µl)	0.2 µl	1.4 µl
DNA extract from original sample	1 µl	
Total volume per reaction	25 µl	

To analyse multiple samples, the reaction mix should be prepared in a batch (in the quantity required for all analyses). To adjust for volume variations during pipetting, it is recommended to increase the number of reactions (n) by 1 for each chip (=n+1), e.g. prepare a reaction mix volume for 7 amplification reactions if 6 samples are to be tested (see Table 3). Always use one vial of MasterMix for the reactions of one chip. Use a precise low volume (0.5-10 µl) pipette for pipetting the Taq DNA Polymerase.



It is recommended to include a negative PCR control for every batch of CarnoCheck® PCR reaction mix prepared. As negative PCR control, the DNA elution buffer of the appropriate DNA extraction kit or PCR-grade water may be used.

- ➡ Mix the reaction mix thoroughly by either vortexing for 2 seconds and then spinning down or by pipetting up and down several times.
- ➡ Aliquot the reaction mix by pipetting **24 µl** of the reaction mix for each PCR reaction into a 0.2 ml, thin-walled PCR reaction tube.



Carry out addition of the template DNA in a separate work space than the set-up of the reaction mix (see Chapter 7.2).

- ➡ Add **1 µl** of DNA extract (5-15 ng/µl) to each PCR reaction and mix either by vortexing for 2 seconds and then spinning down or by pipetting up and down several times. The total volume of one PCR reaction is **25 µl**.
- ➡ Place the reaction tubes in the thermal cycler and start the PCR reaction using the thermal cycler program described in Chapter 8.2.1 (**Table 2**).



After the PCR has been completed, the amplification products should be used immediately for hybridisation or stored in the dark at -20 °C for one week. If so, the PCR products must be heated for 3 minutes at 95 °C before hybridisation.

## 8.3 Hybridisation and washing

### 8.3.1 Preparation and set-up

Hybridisation must be performed at room temperature (20-25 °C). Begin with the necessary preparations for the hybridisation and washing steps at least 30 minutes in advance of starting the hybridisation procedure.

- ▢▢▢▢▢ To dissolve potential precipitates in the hybridisation and washing buffers, expose them to room temperature (20-25 °C) for 30 minutes and mix well before use.

Storage of the **CarnoCheck**® test kit at 4-8 °C may result in precipitates in the Hybridisation Buffer and Buffer B. Allow the solutions to equilibrate to room temperature and then vortex the tube or agitate the bottle until any precipitate is dissolved.

- ▢▢▢▢▢ Prepare the **oCheck**® Hybridisation Chamber: Put a fresh wet paper towel into the Hybridisation Chamber and close the lid to create a humidity-saturated atmosphere

To avoid evaporation of the small volume of used hybridisation mix on the chip, it is necessary to perform the hybridisation in a humidity-saturated atmosphere. A dedicated Hybridisation Chamber for **CarnoCheck**® analysis is available from Greiner Bio-One (see Chapter 2).

- ▢▢▢▢▢ Incubate the required amount of **CarnoCheck**® chips in the prepared Hybridisation Chamber at room temperature (20-25 °C) for at least 10 minutes.



The magnetic slide holder of the Hybridisation Chamber contains a magnet only at one of two ends. If less than four **CarnoCheck**® chips are to be hybridised in parallel, take care to fill the slideholder with **CarnoCheck**® chips from the opposite side of the magnet. Otherwise, the **CarnoCheck**® chips will not be covered with liquid during the washing procedure.

- ▢▢▢▢▢ Prepare the washing solutions I and II appropriate for the number of **CarnoCheck**® chips being analysed. If more than one **CarnoCheck**® chip is processed, multiply the needed volumes for one chip by the number of processed chips (see Table 4).

## Preparation of washing solutions I and II for one CarnoCheck® chip:

- ➡ Washing solution I: Mix 60 ml double-distilled water with 6 ml Buffer A and 750 µl Buffer B and fill it into the WashBox.
- ➡ Washing solution II: Mix 73,5 ml double-distilled water with 900 µl Buffer A and 600 µl Buffer B and fill it into the other WashBox.

washing solution I				
Number of CarnoCheck® chips	1	2	3	4
Distilled/deionised water	60 ml	120 ml	180 ml	240 ml
CarnoCheck® Buffer A	6 ml	12 ml	18 ml	24 ml
CarnoCheck® Buffer B	0.75 ml	1.5 ml	2.25 ml	3 ml
total volume	66.75 ml	133.5 ml	200.25 ml	267 ml

washing solution II				
Number of CarnoCheck® chips	1	2	3	4
Distilled/deionised water	73.5 ml	147 ml	220.5 ml	294 ml
CarnoCheck® Buffer A	0.9 ml	1.8 ml	2.7 ml	3.6 ml
CarnoCheck® Buffer B	0.6 ml	1.2 ml	1.8 ml	2.4 ml
total volume	75 ml	150 ml	225 ml	300 ml

Label them as washing solution I and II. Each oCheck® WashBox contains an engraved scale, indicating the correct amount of washing solution needed for up to 4 chips. Please use this scale to check the buffer quantity.

- ➡ Preheat washing solution II to 50 °C in a temperature-controlled water bath for at least 20 minutes prior to use. Ensure that the fill level of the water bath equals the fill level of the washing solution II.



Never reuse the washing solutions as this could lead to an accumulation of washed-off PCR product that possibly interferes with CarnoCheck® results. Use fresh washing solutions for each assay. The prepared washing solution mix can be stored up to one week at room temperature. Check if precipitation of SDS has occurred. If so, warm up the washing solution mix until the precipitate is dissolved and equilibrate to room temperature again. Then prepare for the next hybridisation experiment.

### 8.3.2 Hybridisation

Hybridisation must be performed at room temperature (20-25 °C). The principle working steps for hybridising the PCR products of the CarnoCheck® PCR reaction onto the CarnoCheck® chip are shown in **Figure 5**.

▢▢▢▢ Mix the PCR products before use. Briefly spin down.

If PCR products were stored at -20 °C until hybridisation, the PCR products must be heated for 3 minutes at 95 °C, mixed, spin briefly down and then proceed as described.

▢▢▢▢ Vortex the Hybridisation Buffer before use. Briefly spin down.

▢▢▢▢ Mix **24 µl** of the CarnoCheck® Hybridisation Buffer in a fresh reaction tube of an 8x PCR strip with **13 µl** of the PCR product by either vortexing or by pipetting up and down several times.

▢▢▢▢ Briefly spin down.

▢▢▢▢ Transfer **30 µl** of the hybridisation mix into each chip well by using six channels of a multichannel pipette. Avoid air bubble formation!

It is recommended to process six samples in parallel using an 8-channel multipipette and 8x PCR strips (**see Figure 5**). This increases handling efficiency, speed and thereby reduces the risk of evaporation. If more than one slide is to be processed at once, the usage of a multipipette is mandatory in order to achieve the correct hybridisation time.

If possible, hybridise all 6 wells of a chip. In case of processing fewer than 6 samples, leave the unused wells empty. Unused wells on a processed chip cannot be used for future samples.



Handle the chip carefully to avoid spilling of the hybridisation mix. Spilling can lead to cross-contamination of samples and to false positive results.


Incubate the chip for exactly 20 minutes at room temperature (20-25 °C) within the prepared Hybridisation Chamber in a dark, humidity-saturated atmosphere. Be careful not to move the Hybridisation Chamber during the hybridisation.




Never change the incubation time or temperature of the hybridisation reaction, as this may cause a loss of fluorescence signal intensity or an increase in unspecific fluorescence.

Do not expose the hybridised chips to direct sunlight.

- ➡ Prepare a humidity-saturated atmosphere in the Hybridisation Chamber.
- ➡ Incubate the required amount of **CarnoCheck®** chips in the Hybridisation Chamber at room temperature (20-25 °C) (see Chapter 8.3.1)



- ➡ Mix 24 µl of the **CarnoCheck®** Hybridisation Buffer in a 0.2 ml reaction tube of a PCR strip with 13 µl of the PCR product. Mix thoroughly.



- ➡ Transfer 30 µl of the hybridisation mix into each well (A1-A6) of the **CarnoCheck®** chip using a multichannel pipette.



- ➡ Close the Hybridisation Chamber and incubate the **CarnoCheck®** chip for exactly 20 minutes at room temperature (20-25 °C).




Figure 5: Working steps of the hybridisation procedure

### 8.3.3 Washing and drying

Special equipment supplied by Greiner Bio-One enables the parallel washing of up to four CarnoCheck® chips (see Chapter 2). The additional equipment required for processing the CarnoCheck® chips is comprised of two oCheck® WashBoxes and a handle for the magnetic slide holder of the Hybridisation Chamber.


The different working steps are shown in **Figure 6**.

- ▢▢▢▢ Carefully remove the magnetic slideholder containing the hybridised slides from the Hybridisation Chamber.
- ▢▢▢▢ Drop the slideholder containing the slides directly into the oCheck® WashBox containing washing solution I. Ensure that the magnetic side is facing up.
- ▢▢▢▢ Attach the oCheck® Handle to the slideholder and begin the first of three washing steps.
- ▢▢▢▢ Wash the chip at room temperature (20-25 °C) in washing solution I by moving it quickly up and down for **20 seconds**. The arrays must stay covered with washing solution at all times.
- ▢▢▢▢ Wash the chip for **20 seconds** in washing **solution II at 50 °C** by vigorously moving the slide holder up and down.
- ▢▢▢▢ **Immediately**, remove any liquid from the chip surface by centrifugation. If a special microcentrifuge for microarrays is used, centrifuge for **1 minute**. If a centrifuge applicable for 50 ml tubes is used, place every washed CarnoCheck® chip into a 50 ml tube and centrifuge at room temperature for **3 minutes at 500 g**.

The CarnoCheck® chip is now ready for scanning and should be scanned immediately. For cleaning of the oCheck® WashBoxes, rinse several times with water after each completed washing and drying procedure.


### First washing step

- ➡ Carefully remove the magnetic slideholder from the Hybridisation Chamber.
- ➡ Quickly drop the slideholder into the oCheck® Washbox with washing solution I.
- ➡ Attach the oCheck® Handle.
- ➡ Wash the CarnoCheck® chip(s) in washing solution I at **room temperature** for **20 seconds** by moving the slideholder up and down.



### Second washing step

- ➡ Wash the CarnoCheck® chip(s) in washing solution II in a water bath **50 °C** for **20 seconds** by moving the slideholder up and down.



### Drying

- ➡ Immediately remove any liquid from the surface of the CarnoCheck® chips by centrifugation.




Figure 6: Working steps of the washing and drying procedure prior to the analysis of the CarnoCheck® chip with the CheckScanner™ and the CheckReport™ Software.

## 8.4 Scanning and evaluation of the CarnoCheck® chip

Place the CarnoCheck® chip(s) into the CheckScanner™ and proceed with scanning as described in detail in the User Guide of the CheckReport™ Software.

For more detailed information about the installation of the CheckScanner™ and the CheckReport™ Software, as well as computer system requirements, please consult the corresponding Instructions For Use of the CheckScanner™ and the CheckReport™ Software

Whenever analysing data using the CheckReport™ Software, ensure that the version of the CheckReport™ Software installed on your computer matches the version indicated on the currently used CarnoCheck® kit. If the versions do not match, update the CheckReport™ Software. The latest Software version can be downloaded from the Greiner Bio-One website:

[www.gbo.com/bioscience/biochips\\_download](http://www.gbo.com/bioscience/biochips_download)



## 9. TROUBLESHOOTING

If one of the following error messages occurs during chip scanning or if the CarnoCheck® analysis fails due to specific on-chip controls, proceed as described below. Please do not hesitate to contact your local Greiner Bio-One distributor if you have any questions or experience any difficulties while using CarnoCheck®.

PROBLEM and cause	Comments and suggestions
<b>ERROR MESSAGE „COULD NOT READ BARCODE“</b>  <b>Damaged barcode</b>  <b>Chip was not loaded correctly</b>	<p>Check barcode for damage. Enter the barcode manually when the appropriate window appears.</p> <p>Check chip orientation and scan the chip in correct orientation.</p>
<b>ERROR MESSAGE „MISSING SPOTS“, PRINTING CONTROL FAILED OR ORIENTATION CONTROL FAILED</b>  <b>Dust on the chip</b>  <b>Formation of air bubbles during transfer of liquid on the chip</b>	<p>Repeat hybridisation of PCR product(s) on another chip.</p> <p>Repeat hybridisation of PCR product(s) on another chip. Pipette carefully to avoid air bubble formation.</p>
<b>HYBRIDISATION CONTROL FAILED</b>  <b>Incorrect temperature of washing solution II</b>  <b>Incorrect temperature of water bath</b>  <b>Wrong preparation of hybridisation mix</b>	<p>The second washing step must be performed at 50 °C. Ensure that the washing solution II is heated to 50 °C.</p> <p>The second washing step must be performed at 50 °C. Check the temperature of the water bath. Ensure that the water bath is set to a temperature of 50 °C. If necessary, confirm temperature using a thermometer.</p> <p>Repeat preparation of hybridisation mix with the correct volumes and hybridise PCR products on another chip.</p>
<b>PCR CONTROL FAILED</b>  <b>No addition of AmpliTaq® Gold to the MasterMix</b>  <b>Addition of a not proper functioning AmpliTaq® Gold to the MasterMix</b>  <b>Insufficient mixing of reaction mix</b>  <b>PCR inhibitors are present in the sample</b>  <b>Hybridisation was performed without addition of PCR product</b>  <b>Insufficient mixing of Hybridisation Mix</b>	<p>Repeat CarnoCheck® analysis starting with the preparation of the PCR reaction.</p> <p>Repeat CarnoCheck® analysis starting with the preparation of the PCR reaction.</p> <p>Repeat CarnoCheck® analysis starting with the PCR reaction. Take care to mix the reaction mix thoroughly.</p> <p>Repeat DNA extraction and CarnoCheck® analysis.</p> <p>Repeat hybridisation.</p> <p>Repeat hybridisation.</p>

PROBLEM and cause	Comments and suggestions
<b>Problems with the thermal cycler</b>	Check both the performance of the thermal cycler and the correct programming of the thermal cycler (PCR steps, heat ramp, volume). Attention: Use either the thermal cycler GeneAmp® PCR System 9700 (Applied Biosystems) or the Veriti™ 96-Well Thermal Cycler (Applied Biosystems) in combination with CarnoCheck®.
<b>PCR AND/OR SAMPLE CONTROLS HAVE NOT FAILED BUT DISPLAY A SNR VALUE OF 0</b>	This result is rated as valid if the CheckReport™ Software detects at least one animal species in the sample with a signal above a defined threshold. The PCR control may then show low or even absent fluorescence signals due to competition during PCR.

## 10. TECHNICAL ASSISTANCE

Greiner Bio-One employs a technical service department staffed with experienced scientists with extensive practical and theoretical expertise in molecular biology and oCheck® products. If you have any questions or experience any difficulties concerning oCheck® products, please do not hesitate to contact your local Greiner Bio-One distributor.

## 11. PERFORMANCE CHARACTERISTICS OF CARNOCHECK®

The CarnoCheck® was tested for the specificity and sensitivity of the analysis of meat products. As reference material sausages with known composition (see table 1) were obtained from the Federal Research Centre for Nutrition and Food in Germany (BfEL, Kulmbach). Two different grades of preservation were tested: cooked preserves (F < 0.9) and full preserves (F = 3.4).

**Table 4: Composition of reference material sausages**

Sample**		Cattle	Pig	Chicken	Turkey	Sheep	Duck*	Horse
<b>S1</b>	% product	0.05	45.7	2.5	1	0.5	0.25	0
	g product	5	4570	250	100	50	25	0
	% meat	0.1	91.4	5	2	1	0.5	0
<b>S2</b>	% product	0.25	0	45.7	2.5	1	0.5	0.05
	g product	25	0	4570	250	100	50	5
	% meat	0.5	0	91.4	5	2	1	0.1
<b>S3</b>	% product	0.5	0.05	0	45.7	2.5	1	0.25
	g product	50	5	0	4570	250	100	25
	% meat	1	0.1	0	91.4	5	2	0.5
<b>S4</b>	% product	1	0.25	0.05	0	45.7	2.5	0.5
	g product	100	25	5	0	4570	250	50
	% meat	2	0.5	0.1	0	91.4	5	1
<b>S5</b>	% product	2.5	0.5	0.25	0.05	0	45.7	1
	g product	250	50	25	5	0	4570	100
	% meat	5	1	0.5	0.1	0	91.4	2
<b>S6</b>	% product	45.7	1	0.5	0.25	0.05	0	2.5
	g product	4570	100	50	25	5	0	20
	% meat	91.4	2	1	0.5	0.1	0	5
<b>S7</b>	% product	0	2.5	1	0.5	0.25	0.05	45.7
	g product	0	250	100	50	25	5	4570
	% meat	0	5	2	1	0.5	0.1	91.4

% product: percentage of meat fractions in total product;

g product: weight proportion of meat fractions in total product;

% meat: percentage of meat fractions in total meat.

\* Detection of duck species is not the intended use of CarnoCheck®, there is no duck specific probe included, the chicken probe might show cross-hybridisation to some (but not all) duck species

\*\* Composition of the tested sausages: 50% meat, 25% plant oil, 23% ice, 1,5% , nitrit pickling salt, 0,25 % spices (pepper, muscat, cardamon, cinnamon)

### 11.4.1 Procedure


DNA extraction was done using the CTAB-method (Binke et al.,2003):

- Lysis buffer: (pH=8), 20 g/l Cetyltrimethylammonium bromide (CTAB), 1.4 mol/l Sodium chloride, 0.1 mol/l Tris(hydroxymethyl)-aminomethane (Tris), 20 mmol/l Ethylenediaminetetraacetate (EDTA)
- Proteinase K solution: 10 mg/ml
- RNase A solution: 20 mg/ml
- Extraction solution: Chloroform / isoamylalcohol (V/V, 24/1)
- TE buffer: 1 mmol/l Tris-HCl (pH=8), 1 mmol/l EDTA (pH=8)


For lysis 25 to 100 mg of meat or 50 mg of meat product is weighed out in a 1.5 ml reaction tube, and to this 400 µl lysis buffer and 20 µl Proteinase K solution are added and mixed. The mixture is allowed to lyse overnight at 56 °C with intermittent shaking (for 10 s every 10 min). For RNA digestion the sample solution is cooled down to room temperature, and 20 µl RNase A solution is added for incubation for 1 h at 37 °C. For nucleic acid extraction 300 µl of the extraction solution is pipetted into the sample, mixed, and centrifuged (>10000 rpm) for 10 min. For nucleic acid precipitation 300 µl isopropanol is added to a new 1.5 ml reaction tube, and 300 µl of the supernatant is pipetted in and mixed well. The solution is incubated at room temperature for 10 min and then centrifuged 10 min at >10000 rpm. The supernatant is carefully decanted and discarded, and the pellet is washed with 500 µl ethanol (70%) and centrifuged for 5 min at >10000 rpm. The supernatant is discarded and the washing step is repeated with ethanol (100%). The pellet is dried at 70 °C for 10 min and taken up in 100 µl TE buffer (prewarmed to 70 °C). 15ng DNA was used as template for the PCR reaction. PCR, hybridisation and washing was done as described in the CarnoCheck® Instructions for Use.

### 11.4.2 Results

In cooked preserves ( $F < 0.9$ ) all tested animal species could be detected down to a percentage of meat fractions of 0.25 in total product and for pig, cattle, turkey and horse even down to 0.05 (see figure A). In full preserves ( $F = 3.4$ ) all tested animal species could be detected down to a percentage of meat fractions of 0.25 in total product and for pig, cattle and turkey even down to 0.05 (see figure B). In a further analysis the limit of detection (LOD) was explored more precisely. For a detailed overview of the determined LODs for each animal species see table 4. For animal species goat and donkey the limit of detection was only determined in mixtures of raw meat. Here LODs of 0.1 or 0.35 respectively were obtained.



**Figure A: Validation of CarnoCheck® using cooked preserves ( $F < 0.9$ ). Signal to noise ratios (SNR) depending on % of meat fractions from different species in total product.**



**Figure B: Validation of CarnoCheck® using full preserves ( $F = 3.4$ ). Signal to noise ratios (SNR) depending on % of meat fractions from different species in total product.**

**Table 5: Limit of detection (LOD) as percentage of meat fractions in total product**

Species	LOD cooked preserves (F < 0.9) (%)	LOD full preserves (F = 3.4) (%)
<b>Pig (<i>Sus scrofa</i>)</b>	0.05	0.05
<b>Cattle (<i>Bos taurus</i>)</b>	0.05	0.05
<b>Sheep (<i>Ovis aries</i>)</b>	0.13	0.25
<b>Turkey (<i>Meleagris gallopavo</i>)</b>	0.1	0.1
<b>Horse (<i>Equus caballus</i>)</b>	0.05	0.25
<b>Chicken (<i>Gallus gallus</i>)</b>	0.16	0.25
<b>Goat (<i>Capra hircus</i>)</b>	0.1*	-
<b>Donkey (<i>Equus asinus</i>)</b>	0.35*	-

\* Only tested in mixtures of raw meat.

### 11.4.3 Successful participation at FAPAS® Proficiency Test

The FAPAS® (Food Analysis Performance Assessment Scheme, UK) proficiency testing is an essential element of laboratory quality assurance and is an important requirement of the EU Additional Measures Directive 93/99/EEC applying to laboratories entrusted with the official control of food. The analysis of an external quality check sample provides objective standards for quality of the analysis of an individual laboratory. In FAPAS® test 2911 freeze dried, raw chicken test material was dispatched to the participants and was analysed for contaminating meat fractions. These contaminations (beef and pork) were distinctively identified by using the CarnoCheck® test system. The FAPAS® test material for FAPAS® test 2918 was freeze-dried raw pork contaminated with lamb. The test participants were told that it contained at least one contaminant from lamb, chicken or beef. CarnoCheck® determined the presence of lamb and pork and absence of beef and chicken or other contaminants. So CarnoCheck® participated successfully at the FAPAS® Proficiency Test 2911 and 2918.


## 12. CARNOCHECK® SHORT PROTOCOL

### 12.1 Room 2: PCR - Set-up of reaction mix

Sample collection, homogenisation and DNA extraction have to be performed in a separate room and are not part of this short protocol.


- ▶ Prepare the reaction mix for the required quantity of PCR reactions
- ▶ Mix the PCR MasterMix carefully before pipetting

	1 reaction	7 reactions (1 chip)
CarnoCheck® PCR MasterMix	23.8 µl	166.6 µl
AmpliTaq Gold® DNA Polymerase (5 U /µl)	0.2 µl	1.4 µl
<b>Total volume</b>	<b>24 µl</b>	<b>168 µl</b>




- ▶ Mix the reaction mix carefully

- ▶ Aliquot the reaction mix: add **24 µl** of the reaction mix for each PCR reaction into a 0.2 ml PCR reaction tube of a PCR strip




## 12.2 Room 2: PCR - DNA template addition / PCR reaction

- ➡ Add 1  $\mu$ l of DNA template for each PCR reaction
- ➡ Mix thoroughly



- ➡ Start the PCR reaction with the prepared thermal cycler program

Time	Temp. °C	No. of cycles
10 min	94 °C	1
30 s	94 °C	40
30 s	53 °C	
30 s	72 °C	
5 min	72 °C	1
Hold	10 °C	





## 12.3 Room 3: Hybridisation - Preparation / Hybridisation reaction



Begin preparation at least 30 minutes prior to hybridisation.



- Prepare the washing solution I and II for the number of CarnoCheck® chips to be analysed.

washing solution I				
Number of CarnoCheck® chips	1	2	3	4
Distilled/deionised water	60 ml	120 ml	180 ml	240 ml
CarnoCheck® Buffer A	6 ml	12 ml	18 ml	24 ml
CarnoCheck® Buffer B	0.75 ml	1.5 ml	2.25 ml	3 ml
total volume	66.75 ml	133.5 ml	200.25 ml	267 ml
washing solution II				
Number of CarnoCheck® chips	1	2	3	4
Distilled/deionised water	73.5 ml	147 ml	220.5 ml	294 ml
CarnoCheck® Buffer A	0.9 ml	1.8 ml	2.7 ml	3.6 ml
CarnoCheck® Buffer B	0.6 ml	1.2 ml	1.8 ml	2.4 ml
total volume	75 ml	150 ml	225 ml	300 ml
<b>Total volume</b>	<b>155.75 ml</b>	<b>311.50 ml</b>	<b>467.25 ml</b>	<b>623 ml</b>






- Fill washing solution I and II in separate oCheck® WashBoxes
- Preheat washing solution II in a water bath at 50 °C

- Incubate the number of CarnoCheck® chips to be analysed in the prepared oCheck® Hybridisation Chamber at room temperature

- Mix PCR products and briefly spin down
- Mix Hybridisation Buffer and briefly spin down
- Mix **24 µl** CarnoCheck® Hybridisation Buffer with **13 µl** PCR product
- Mix thoroughly and briefly spin down



## 12.4 Room 3: Hybridisation - Preparation / Hybridisation reaction

<ul style="list-style-type: none"> <li>Transfer <b>30 µl</b> of the hybridisation mix into each well of the CarnoCheck® chip using a multichannel pipette</li> <li>Avoid air bubble formation</li> </ul>	
<ul style="list-style-type: none"> <li>Incubate the CarnoCheck® chip for exactly <b>20 minutes</b> at <b>room temperature (20-25 °C)</b></li> </ul>	
<ul style="list-style-type: none"> <li>Remove the magnetic slideholder from the Hybridisation Chamber</li> <li>Drop the slideholder into the oCheck® Washbox with washing solution I</li> <li>Attach the oCheck® Handle</li> <li>Wash the CarnoCheck® chip(s) in washing solution I at <b>room temperature</b> for <b>20 seconds</b></li> </ul>	
<ul style="list-style-type: none"> <li>Wash the CarnoCheck® chip(s) in preheated washing solution II in a water bath at <b>50°C</b> for <b>20 seconds</b></li> </ul>	
<ul style="list-style-type: none"> <li>Remove any liquid from the CarnoCheck® chip surface by centrifugation</li> </ul>	 <p><b>3 minutes</b> <b>500 g</b> or <b>1 minute</b> <b>max. speed</b></p>
<ul style="list-style-type: none"> <li>Scan the CarnoCheck® chip(s) with the CheckScanner™</li> <li>Perform scanning and analysis as described in the User Guide of the CheckReport™ Software</li> <li>Create reports</li> </ul>	